

SPONSOR 13714

George Milner
Fluoro-Plastics
3601 G Street
Philadelphia, PA 19134

REPORT DATE: 3/2/15



PAGE: 1 of 1

CERTIFICATE
CLASS VI - 121°C PLASTIC TESTS
(USP 37 <88>, REV. 5/2014)

Name: Skived Virgin PTFE sheet/film
Physical Description: Solid
Total Quantity Received for Testing: 1 sheet
Lot Number: 14242
Sample Code: Dupont 6270
Part Number: VS003X12A
Storage Conditions: Room Temperature
Date Received: 1/7/15
Class VI - Intracutaneous Test
Pacific BioLabs Report No.: 15A0035H-X02G

Class VI - Systemic Injection Test
Pacific BioLabs Report No.: 15A0035H-X03G

Class VI - Intramuscular Implantation Test
Pacific BioLabs Report No.: 15A0035H-X04G

The test article met the requirements for Plastic Class VI - 121°C testing per USP <88>.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist
3/3/15

X 

I approve the content of this document.
Signed by: Zuzana Karjala

3/2/15

X 

QA Review
Signed by: Christopher Juan
:sdb